

Equipment Inventory

Name:	Brand name	Description:	Known Issues:	Lab:	Pic
Digital Oscilloscope	Tektronix TPS 2024	200 MHZ bandwidth, 2 GS/s, four channels and an isolated external trigger			 126
Centrifuge	Jouan B4i	Max rpm: 4300, s40, Max load: 2.00 Kg, 120 volts, 60 HZ, Five acceleration and Five braking rates			 126
Orbital Rotator	Thermolyne-Rotomix	Speed Range :10-60, orbit size (cm): 2, Ambient Conditions °c: 4-37, Max Load (kg):2.2			 126
Vortex Mixer	Maxi Mix II- Type 37600 Mixer	Speed (rpm): 100-3000			 126
(Heats and Stirs)???	Cole Parmer				 126
pH Tester	Oakton-pHTestr BNC 10	waterproof, pH range: -1.00 to 15.00, Accuracy: ± .01, Operating temperature: 32 to 122°F, calibration: up to 3			 126
Balance	Denver Instrument XS-310D	Max1: 31 g, Max2: 310 g			 126
Chilling Incubator	Echo Therm in30-Torrey Pines Scientific	Temp range: 4-70 °C ,display can be changed from temperature to timer, Max set time: 99 Hrs, Accuracy: ± 0.2 °C			 126
Digital Incubator	Boekel Scientific-Model 133000	Heat: Knobs 0-10, 115 volt/50-60 Hz, Temp range: +5°C to 60°C			 126
					126
					126